

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
 Data File : VN078280.D
 Acq On : 19 Jun 2023 15:24
 Operator : JC\MD
 Sample : 03310-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW11-GW-883-885

Quant Time: Jun 20 01:28:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	470496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	874155	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	785149	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	247520	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	333580	53.176	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.360%
35) Dibromofluoromethane	8.177	113	303304	54.875	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.760%
50) Toluene-d8	10.576	98	1002503	47.049	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.100%
62) 4-Bromofluorobenzene	12.853	95	330200	49.849	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.700%
Target Compounds						
16) Acetone	4.459	43	5155	2.855	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
Data File : VN078280.D
Acq On : 19 Jun 2023 15:24
Operator : JC\MD
Sample : 03310-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW11-GW-883-885

Quant Time: Jun 20 01:28:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

