

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
 Data File : VN078281.D
 Acq On : 19 Jun 2023 15:48
 Operator : JC\MD
 Sample : 03310-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-RW5-TB-20230609

Quant Time: Jun 20 01:28:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	459717	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	851884	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	746110	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	220735	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	329471	53.752	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.500%
35) Dibromofluoromethane	8.171	113	296086	54.970	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.940%
50) Toluene-d8	10.576	98	976758	47.039	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.080%
62) 4-Bromofluorobenzene	12.853	95	301381	46.687	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.380%
Target Compounds						
16) Acetone	4.465	43	2888	1.637	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
Data File : VN078281.D
Acq On : 19 Jun 2023 15:48
Operator : JC\MD
Sample : 03310-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-RW5-TB-20230609

Quant Time: Jun 20 01:28:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

