

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
 Data File : VN078286.D
 Acq On : 19 Jun 2023 18:13
 Operator : JC\MD
 Sample : PB153531TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB153531TB

Quant Time: Jun 20 01:31:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	452155	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	847325	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	778835	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	239586	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	322180	53.442	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.880%			
35) Dibromofluoromethane	8.177	113	290151	54.158	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 108.320%			
50) Toluene-d8	10.571	98	994333	48.143	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 96.280%			
62) 4-Bromofluorobenzene	12.853	95	320613	49.934	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 99.860%			
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	155871	89.826	ug/l	98
20) Methylene Chloride	5.289	84	16597	2.894	ug/l	94
25) 2-Butanone	7.494	43	53706	20.126	ug/l #	83
37) Ethyl Acetate	7.577	43	79501	13.243	ug/l	97
43) Isopropyl Acetate	8.700	43	348191	30.327	ug/l #	82
51) 4-Methyl-2-Pentanone	10.453	43	351946	58.213	ug/l	91
89) n-Butylbenzene	14.059	91	13401	1.003	ug/l	97
93) 1,2,4-Trichlorobenzene	15.400	180	4389	1.417	ug/l	96
95) Naphthalene	15.647	128	15960	1.532	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
Data File : VN078286.D
Acq On : 19 Jun 2023 18:13
Operator : JC\MD
Sample : PB153531TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB153531TB

Quant Time: Jun 20 01:31:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

