

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
 Data File : VN078287.D
 Acq On : 19 Jun 2023 18:38
 Operator : JC\MD
 Sample : 03320-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-19

Quant Time: Jun 20 01:32:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	455383	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	863102	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	807469	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	254450	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	329874	54.330	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.660%		
35) Dibromofluoromethane	8.177	113	296719	54.371	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 108.740%		
50) Toluene-d8	10.576	98	1025488	48.744	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 97.480%		
62) 4-Bromofluorobenzene	12.853	95	341628	52.234	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 104.460%		
Target Compounds						
16) Acetone	4.447	43	27631	15.810	ug/l	95
20) Methylene Chloride	5.283	84	16034	2.776	ug/l #	81
37) Ethyl Acetate	7.577	43	63843	10.440	ug/l	97
43) Isopropyl Acetate	8.700	43	346357	29.616	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
Data File : VN078287.D
Acq On : 19 Jun 2023 18:38
Operator : JC\MD
Sample : 03320-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-19

Quant Time: Jun 20 01:32:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

