

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
 Data File : VN078292.D
 Acq On : 19 Jun 2023 20:39
 Operator : JC\MD
 Sample : 03320-17
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-16

Quant Time: Jun 20 01:35:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	453989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	860996	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	808241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	258646	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	322945	53.352	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.700%			
35) Dibromofluoromethane	8.177	113	298623	54.854	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 109.700%			
50) Toluene-d8	10.576	98	1020984	48.648	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.300%			
62) 4-Bromofluorobenzene	12.859	95	335501	51.423	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 102.840%			
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	25832	14.826	ug/l #	81
20) Methylene Chloride	5.289	84	15104	2.623	ug/l #	82
37) Ethyl Acetate	7.577	43	63029	10.332	ug/l	96
43) Isopropyl Acetate	8.694	43	563131	48.270	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
Data File : VN078292.D
Acq On : 19 Jun 2023 20:39
Operator : JC\MD
Sample : 03320-17
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-16

Quant Time: Jun 20 01:35:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

