

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
 Data File : VN078293.D
 Acq On : 19 Jun 2023 21:03
 Operator : JC\MD
 Sample : 03320-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14A

Quant Time: Jun 20 01:36:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	458593	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	876987	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	791621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	244617	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	329205	53.841	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.680%	
35) Dibromofluoromethane	8.177	113	301327	54.341	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.680%	
50) Toluene-d8	10.577	98	1013397	47.406	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 94.820%	
62) 4-Bromofluorobenzene	12.853	95	325179	48.932	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.860%	
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	26428	15.016	ug/l #	67
20) Methylene Chloride	5.295	84	16288	2.801	ug/l #	80
37) Ethyl Acetate	7.577	43	66550	10.711	ug/l #	97
43) Isopropyl Acetate	8.700	43	465101	39.140	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
Data File : VN078293.D
Acq On : 19 Jun 2023 21:03
Operator : JC\MD
Sample : 03320-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14A

Quant Time: Jun 20 01:36:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

