

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
 Data File : VN078294.D
 Acq On : 19 Jun 2023 21:27
 Operator : JC\MD
 Sample : 03321-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-25

Quant Time: Jun 20 01:36:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	469396	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	883133	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	820196	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	272615	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	330746	52.848	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.700%
35) Dibromofluoromethane	8.177	113	294235	52.693	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.380%
50) Toluene-d8	10.571	98	1041569	48.385	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.780%
62) 4-Bromofluorobenzene	12.853	95	346108	51.719	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.440%
Target Compounds						
						Qvalue
16) Acetone	4.447	43	24333	13.508	ug/l	91
20) Methylene Chloride	5.289	84	16618	2.792	ug/l #	87
37) Ethyl Acetate	7.577	43	59978	9.586	ug/l	97
43) Isopropyl Acetate	8.700	43	564773	47.197	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061923\
Data File : VN078294.D
Acq On : 19 Jun 2023 21:27
Operator : JC\MD
Sample : 03321-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-25

Quant Time: Jun 20 01:36:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

