

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
 Data File : VN082632.D
 Acq On : 19 Jun 2024 11:21
 Operator : JC\MD
 Sample : VN0619WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jun 20 02:56:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	183931	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	303189	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	279200	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	130790	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	153641	52.089	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.180%	
35) Dibromofluoromethane	8.171	113	107196	52.118	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.240%	
50) Toluene-d8	10.571	98	410942	52.833	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.660%	
62) 4-Bromofluorobenzene	12.853	95	134835	50.281	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 100.560%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	48753	18.979	ug/l	99
3) Chloromethane	2.359	50	53215	17.281	ug/l	94
4) Vinyl Chloride	2.512	62	52525	17.923	ug/l	98
5) Bromomethane	2.954	94	33294	17.649	ug/l	100
6) Chloroethane	3.118	64	36376	17.407	ug/l	99
7) Trichlorofluoromethane	3.495	101	68787	17.991	ug/l	98
8) Diethyl Ether	3.959	74	23278	17.877	ug/l	99
9) 1,1,2-Trichlorotrifluo...	4.377	101	40013	18.684	ug/l	98
10) Methyl Iodide	4.595	142	37097	18.185	ug/l	94
11) Tert butyl alcohol	5.536	59	44994	93.205	ug/l	99
12) 1,1-Dichloroethene	4.342	96	35705	17.435	ug/l	90
13) Acrolein	4.177	56	40605	82.571	ug/l	97
14) Allyl chloride	5.024	41	77283	19.296	ug/l	89
15) Acrylonitrile	5.724	53	118727	93.173	ug/l	99
16) Acetone	4.436	43	117727	87.762	ug/l	100
17) Carbon Disulfide	4.712	76	109310	16.807	ug/l	97
18) Methyl Acetate	5.030	43	61382	18.709	ug/l	100
19) Methyl tert-butyl Ether	5.800	73	130118	18.562	ug/l	100
20) Methylene Chloride	5.277	84	43807	18.575	ug/l	96
21) trans-1,2-Dichloroethene	5.789	96	38651	17.758	ug/l	97
22) Diisopropyl ether	6.677	45	156965	18.687	ug/l	97
23) Vinyl Acetate	6.606	43	607660	91.229	ug/l	100
24) 1,1-Dichloroethane	6.571	63	84778	18.632	ug/l	99
25) 2-Butanone	7.489	43	177453	93.069	ug/l	98
26) 2,2-Dichloropropane	7.489	77	72212	18.294	ug/l	100
27) cis-1,2-Dichloroethene	7.494	96	45912	17.507	ug/l	98
28) Bromochloromethane	7.818	49	41308	19.893	ug/l	99
29) Tetrahydrofuran	7.841	42	116546	99.849	ug/l	100
30) Chloroform	7.971	83	83589	18.833	ug/l	95
31) Cyclohexane	8.259	56	75233	17.674	ug/l	94
32) 1,1,1-Trichloroethane	8.171	97	73613	18.289	ug/l	100
36) 1,1-Dichloropropene	8.371	75	57113	18.575	ug/l	100
37) Ethyl Acetate	7.565	43	69652	17.784	ug/l	99
38) Carbon Tetrachloride	8.365	117	64139	18.510	ug/l	93
39) Methylcyclohexane	9.606	83	57091	17.644	ug/l	95
40) Benzene	8.612	78	178137	18.750	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
 Data File : VN082632.D
 Acq On : 19 Jun 2024 11:21
 Operator : JC\MD
 Sample : VN0619WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Jun 20 02:56:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	39464	19.712	ug/l	98
42) 1,2-Dichloroethane	8.677	62	67914	18.793	ug/l	99
43) Isopropyl Acetate	8.694	43	123408	16.860	ug/l	99
44) Trichloroethene	9.353	130	42385	19.213	ug/l	89
45) 1,2-Dichloropropane	9.624	63	46039	18.827	ug/l	93
46) Dibromomethane	9.706	93	30875	18.626	ug/l	96
47) Bromodichloromethane	9.888	83	67811	19.187	ug/l	99
48) Methyl methacrylate	9.682	41	56033	18.570	ug/l	99
49) 1,4-Dioxane	9.700	88	19209	406.992	ug/l	97
51) 4-Methyl-2-Pentanone	10.447	43	370262	98.625	ug/l	99
52) Toluene	10.635	92	109081	18.946	ug/l	98
53) t-1,3-Dichloropropene	10.841	75	68371	18.924	ug/l	99
54) cis-1,3-Dichloropropene	10.318	75	72483	19.401	ug/l	97
55) 1,1,2-Trichloroethane	11.018	97	40880	19.617	ug/l	95
56) Ethyl methacrylate	10.877	69	64337	18.641	ug/l	98
57) 1,3-Dichloropropane	11.165	76	74642	19.375	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.165	63	163151	89.358	ug/l	99
59) 2-Hexanone	11.200	43	280255	98.218	ug/l	98
60) Dibromochloromethane	11.365	129	49246	19.359	ug/l	100
61) 1,2-Dibromoethane	11.471	107	42620	19.398	ug/l	99
64) Tetrachloroethene	11.106	164	40131	17.433	ug/l	97
65) Chlorobenzene	11.894	112	117133	18.295	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.965	131	42346	18.548	ug/l	99
67) Ethyl Benzene	11.965	91	196333	17.997	ug/l	100
68) m/p-Xylenes	12.076	106	151126	37.236	ug/l	98
69) o-Xylene	12.400	106	70366	18.454	ug/l	98
70) Styrene	12.412	104	122407	18.914	ug/l	99
71) Bromoform	12.582	173	32400	17.950	ug/l #	100
73) Isopropylbenzene	12.700	105	181393	18.264	ug/l	99
74) N-amyl acetate	12.494	43	96498	18.206	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.941	83	60714	18.741	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	83013	25.950	ug/l	40
77) Bromobenzene	12.982	156	46660	18.252	ug/l	100
78) n-propylbenzene	13.035	91	215018	18.303	ug/l	99
79) 2-Chlorotoluene	13.129	91	133535	18.126	ug/l	100
80) 1,3,5-Trimethylbenzene	13.176	105	150795	18.710	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.741	75	22660	16.947	ug/l	89
82) 4-Chlorotoluene	13.223	91	133881	18.192	ug/l	99
83) tert-Butylbenzene	13.441	119	125358	18.260	ug/l	99
84) 1,2,4-Trimethylbenzene	13.482	105	152648	18.880	ug/l	99
85) sec-Butylbenzene	13.618	105	176874	18.140	ug/l	98
86) p-Isopropyltoluene	13.729	119	146428	18.649	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	83493	18.013	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	83206	17.174	ug/l	99
89) n-Butylbenzene	14.059	91	120579	17.313	ug/l	99
90) Hexachloroethane	14.335	117	31816	18.624	ug/l	98
91) 1,2-Dichlorobenzene	14.106	146	80025	17.682	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.723	75	12946	18.882	ug/l	97
93) 1,2,4-Trichlorobenzene	15.394	180	39365	16.495	ug/l	99
94) Hexachlorobutadiene	15.506	225	17704	15.046	ug/l	97
95) Naphthalene	15.641	128	126206	17.189	ug/l	100
96) 1,2,3-Trichlorobenzene	15.847	180	42401	17.434	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
Data File : VN082632.D
Acq On : 19 Jun 2024 11:21
Operator : JC\MD
Sample : VN0619WBS01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :


Quant Time: Jun 20 02:56:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 20 02:56:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.230 min Scan# 10

Delta R.T. 0.001 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:168 Resp: 183931

Ion Ratio Lower Upper

168 100

99 58.0 48.2 72.4



#2

Dichlorodifluoromethane

Concen: 18.979 ug/l

RT: 2.124 min Scan# 29

Delta R.T. 0.000 min

Lab File: VN082632.D




Acq: 19 Jun 2024 11:21


Tgt Ion: 85 Resp: 48753

Ion Ratio Lower Upper

85 100

87 30.5 15.7 47.0





#3

Chloromethane

Concen: 17.281 ug/l

RT: 2.359 min Scan# 69

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 50 Resp: 53215

Ion Ratio Lower Upper

50 100

52 34.4 25.0 37.4



#4

Vinyl Chloride

Concen: 17.923 ug/l

RT: 2.512 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN082632.D




Acq: 19 Jun 2024 11:21


Tgt Ion: 62 Resp: 52525

Ion Ratio Lower Upper

62 100

64 30.3 25.0 37.6









#5
Bromomethane
Concen: 17.649 ug/l
RT: 2.954 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 94 Resp: 33294
Ion Ratio Lower Upper
94 100
96 97.1 78.0 117.0



#6
Chloroethane
Concen: 17.407 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.006 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21


Tgt Ion: 64 Resp: 36376
Ion Ratio Lower Upper
64 100
66 32.8 26.6 40.0








#7
 Trichlorofluoromethane
 Concen: 17.991 ug/l
 RT: 3.495 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: VN082632.D
 Acq: 19 Jun 2024 11:21

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion:101 Resp: 68787
 Ion Ratio Lower Upper
 101 100
 103 64.4 52.7 79.1



#8
 Diethyl Ether
 Concen: 17.877 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. -0.006 min
 Lab File: VN082632.D
 Acq: 19 Jun 2024 11:21

Tgt Ion: 74 Resp: 23278
 Ion Ratio Lower Upper
 74 100
 45 110.0 55.5 166.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.684 ug/l
RT: 4.377 min Scan# 412
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion:101 Resp: 40013
Ion Ratio Lower Upper
101 100
85 46.0 38.3 57.5
151 73.3 59.3 88.9



#10
Methyl Iodide
Concen: 18.185 ug/l
RT: 4.595 min Scan# 449
Delta R.T. 0.006 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21


Tgt Ion:142 Resp: 37097
Ion Ratio Lower Upper
142 100
127 42.8 37.5 56.3
141 13.9 13.1 19.7








#11
Tert butyl alcohol
Concen: 93.205 ug/l
RT: 5.536 min Scan# 609
Delta R.T. 0.012 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion: 59 Resp: 44994
Ion Ratio Lower Upper
59 100
57 11.5 8.9 13.3



#12
1,1-Dichloroethene
Concen: 17.435 ug/l
RT: 4.342 min Scan# 406
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21



Tgt Ion: 96 Resp: 35705
Ion Ratio Lower Upper
96 100
61 169.0 149.7 224.5
98 61.0 50.1 75.1





#13

Acrolein

Concen: 82.571 ug/l

RT: 4.177 min Scan# 31

Delta R.T. 0.000 min


Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 56 Resp: 40605

Ion Ratio Lower Upper

56 100

55 73.0 56.4 84.6



#14

Allyl chloride

Concen: 19.296 ug/l

RT: 5.024 min Scan# 522

Delta R.T. 0.006 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21




Tgt Ion: 41 Resp: 77283


Ion Ratio Lower Upper

41 100

39 73.4 67.8 101.6


76 28.3 25.8 38.8








#15
Acrylonitrile
Concen: 93.173 ug/l
RT: 5.724 min Scan# 641
Delta R.T. 0.006 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Instrument :
MSVOA_N
ClientSampleId :



Tgt Ion: 53 Resp: 118727




Ion	Ratio	Lower	Upper
53	100		
52	83.7	66.6	100.0
51	38.8	30.1	45.1




#16
Acetone
Concen: 87.762 ug/l
RT: 4.436 min Scan# 422
Delta R.T. 0.006 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion: 43 Resp: 117727

Ion	Ratio	Lower	Upper
43	100		
58	27.0	21.8	32.6








#17
Carbon Disulfide
Concen: 16.807 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 109310
Ion Ratio Lower Upper
76 100
78 10.2 7.2 10.8



#18
Methyl Acetate
Concen: 18.709 ug/l
RT: 5.030 min Scan# 523
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion: 43 Resp: 61382
Ion Ratio Lower Upper
43 100
74 20.6 16.3 24.5





#19

Methyl tert-butyl Ether

Concen: 18.562 ug/l

RT: 5.800 min Scan# 61

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 73 Resp: 130118

Ion Ratio Lower Upper

73 100

57 22.2 17.9 26.9



#20

Methylene Chloride

Concen: 18.575 ug/l

RT: 5.277 min Scan# 565

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Tgt Ion: 84 Resp: 43807




Ion Ratio Lower Upper

84 100

49 142.0 119.6 179.4

51 44.7 34.8 52.2


86 68.2 52.9 79.3








#21
trans-1,2-Dichloroethene
Concen: 17.758 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 96 Resp: 38651
Ion Ratio Lower Upper
96 100
61 152.7 123.4 185.2
98 65.9 49.3 73.9



#22
Diisopropyl ether
Concen: 18.687 ug/l
RT: 6.677 min Scan# 803
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21


Tgt Ion: 45 Resp: 156965
Ion Ratio Lower Upper
45 100
43 58.6 44.0 66.0
87 24.3 19.7 29.5
59 11.5 8.5 12.7







#23
 Vinyl Acetate
 Concen: 91.229 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.000 min
 Lab File: VN082632.D
 Acq: 19 Jun 2024 11:21

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion: 43 Resp: 607660
 Ion Ratio Lower Upper
 43 100
 86 9.0 7.3 10.9



#24
 1,1-Dichloroethane
 Concen: 18.632 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. 0.000 min
 Lab File: VN082632.D
 Acq: 19 Jun 2024 11:21

Tgt Ion: 63 Resp: 84778
 Ion Ratio Lower Upper
 63 100
 98 6.8 3.3 9.9
 100 4.5 2.0 6.0









#25
2-Butanone
Concen: 93.069 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 177453
Ion Ratio Lower Upper
43 100
72 23.5 18.2 27.2



#26
2,2-Dichloropropane
Concen: 18.294 ug/l
RT: 7.489 min Scan# 941
Delta R.T. -0.006 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion: 77 Resp: 72212
Ion Ratio Lower Upper
77 100
97 20.8 10.3 30.9









#27
 cis-1,2-Dichloroethene
 Concen: 17.507 ug/l
 RT: 7.494 min Scan# 941
 Delta R.T. 0.006 min
 Lab File: VN082632.D
 Acq: 19 Jun 2024 11:21




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 96 Resp: 45912
 Ion Ratio Lower Upper
 96 100
 61 168.5 0.0 329.6
 98 65.6 0.0 128.6



#28
 Bromochloromethane
 Concen: 19.893 ug/l
 RT: 7.818 min Scan# 997
 Delta R.T. 0.006 min
 Lab File: VN082632.D
 Acq: 19 Jun 2024 11:21

Tgt Ion: 49 Resp: 41308
 Ion Ratio Lower Upper
 49 100
 129 1.2 0.0 3.8
 130 63.6 50.5 75.7





#29

Tetrahydrofuran

Concen: 99.849 ug/l

RT: 7.841 min Scan# 1001

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 42 Resp: 116546

Ion Ratio Lower Upper

42 100

72 37.7 30.0 45.0

71 35.5 28.3 42.5



#30

Chloroform

Concen: 18.833 ug/l

RT: 7.971 min Scan# 1023

Delta R.T. 0.006 min

Lab File: VN082632.D




Acq: 19 Jun 2024 11:21


Tgt Ion: 83 Resp: 83589

Ion Ratio Lower Upper

83 100


85 67.3 50.9 76.3








#31
Cyclohexane
Concen: 17.674 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 56 Resp: 75233
Ion Ratio Lower Upper
56 100
69 28.4 24.5 36.7
84 76.9 66.0 99.0



#32
1,1,1-Trichloroethane
Concen: 18.289 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion: 97 Resp: 73613
Ion Ratio Lower Upper
97 100
99 65.0 52.0 78.0
61 52.7 42.1 63.1





#33

1,2-Dichloroethane-d4

Concen: 52.089 ug/l

RT: 8.583 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 65 Resp: 153641

Ion Ratio Lower Upper

65 100

67 50.4 0.0 104.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21




Tgt Ion: 114 Resp: 303189


Ion Ratio Lower Upper

114 100

63 21.9 0.0 44.6

88 16.2 0.0 31.4





#35

Dibromofluoromethane

Concen: 52.118 ug/l

RT: 8.171 min Scan# 1056

Delta R.T. 0.006 min


Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion:113 Resp: 107196

Ion Ratio Lower Upper

113 100

111 103.8 82.4 123.6

192 18.8 14.9 22.3



#36

1,1-Dichloropropene

Concen: 18.575 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21




Tgt Ion: 75 Resp: 57113


Ion Ratio Lower Upper

75 100

110 32.8 16.3 48.9

77 30.8 24.6 37.0









#37
Ethyl Acetate
Concen: 17.784 ug/l
RT: 7.565 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 69652
Ion Ratio Lower Upper
43 100
61 12.1 9.3 13.9
70 9.2 7.0 10.6



#38
Carbon Tetrachloride
Concen: 18.510 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion: 117 Resp: 64139
Ion Ratio Lower Upper
117 100
119 102.4 74.9 112.3
121 29.6 24.3 36.5









#39
Methylcyclohexane
Concen: 17.644 ug/l
RT: 9.606 min Scan# 11301
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 83 Resp: 57091
Ion Ratio Lower Upper
83 100
55 78.7 66.4 99.6
98 47.7 40.7 61.1



#40
Benzene
Concen: 18.750 ug/l
RT: 8.612 min Scan# 1132
Delta R.T. 0.006 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion: 78 Resp: 178137
Ion Ratio Lower Upper
78 100
77 23.8 19.0 28.4








#41
Methacrylonitrile
Concen: 19.712 ug/l
RT: 7.783 min Scan# 991
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 41	Resp: 39464
Ion Ratio	Lower Upper
41 100	
39 57.9	45.5 68.3
67 61.0	50.5 75.7
52 29.8	24.0 36.0



#42
1,2-Dichloroethane
Concen: 18.793 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. 0.006 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21


Tgt Ion: 62	Resp: 67914
Ion Ratio	Lower Upper
62 100	
98 8.3	0.0 15.8








#43
Isopropyl Acetate
Concen: 16.860 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 43 Resp: 123408
Ion Ratio Lower Upper
43 100
61 22.7 17.8 26.6
87 10.9 8.2 12.2



#44
Trichloroethene
Concen: 19.213 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion:130 Resp: 42385
Ion Ratio Lower Upper
130 100
95 87.8 0.0 197.8





#45

1,2-Dichloropropane

Concen: 18.827 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 63 Resp: 46039

Ion Ratio Lower Upper

63 100

65 34.2 24.4 36.6



#46

Dibromomethane

Concen: 18.626 ug/l

RT: 9.706 min Scan# 1318

Delta R.T. -0.006 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21




Tgt Ion: 93 Resp: 30875


Ion Ratio Lower Upper

93 100

95 84.2 65.8 98.6

174 94.0 71.7 107.5









#47
Bromodichloromethane
Concen: 19.187 ug/l
RT: 9.888 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 67811
Ion Ratio Lower Upper
83 100
85 62.7 51.1 76.7
127 9.0 7.2 10.8



#48
Methyl methacrylate
Concen: 18.570 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion: 41 Resp: 56033
Ion Ratio Lower Upper
41 100
69 69.7 56.3 84.5
39 63.8 50.3 75.5





#49

1,4-Dioxane

Concen: 406.992 ug/l

RT: 9.700 min Scan# 11

Delta R.T. 0.006 min


Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N

ClientSampleId :






Tgt Ion: 88 Resp: 19209

Ion Ratio Lower Upper

88 100

43 34.2 27.8 41.8

58 77.7 59.4 89.0



#50

Toluene-d8

Concen: 52.833 ug/l

RT: 10.571 min Scan# 1465

Delta R.T. 0.000 min

Lab File: VN082632.D




Acq: 19 Jun 2024 11:21


Tgt Ion: 98 Resp: 410942

Ion Ratio Lower Upper

98 100

100 64.1 51.5 77.3





#51

4-Methyl-2-Pentanone

Concen: 98.625 ug/l

RT: 10.447 min Scan# 1444

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 43 Resp: 370262

Ion Ratio Lower Upper

43 100

58 34.7 28.2 42.4



#52

Toluene

Concen: 18.946 ug/l

RT: 10.635 min Scan# 1476

Delta R.T. 0.000 min

Lab File: VN082632.D




Acq: 19 Jun 2024 11:21


Tgt Ion: 92 Resp: 109081

Ion Ratio Lower Upper

92 100

91 176.3 139.4 209.0









#53
 t-1,3-Dichloropropene
 Concen: 18.924 ug/l
 RT: 10.841 min Scan# 1511
 Delta R.T. 0.006 min
 Lab File: VN082632.D
 Acq: 19 Jun 2024 11:21




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 75 Resp: 68371
 Ion Ratio Lower Upper
 75 100
 77 32.9 25.7 38.5



#54
 cis-1,3-Dichloropropene
 Concen: 19.401 ug/l
 RT: 10.318 min Scan# 1422
 Delta R.T. 0.000 min
 Lab File: VN082632.D
 Acq: 19 Jun 2024 11:21

Tgt Ion: 75 Resp: 72483
 Ion Ratio Lower Upper
 75 100
 77 31.7 24.3 36.5
 39 60.4 50.5 75.7





#55

1,1,2-Trichloroethane

Concen: 19.617 ug/l

RT: 11.018 min Scan# 1541

Delta R.T. 0.000 min


Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion: 97 Resp: 40880




Ion Ratio Lower Upper

97 100

83 95.1 71.4 107.0

85 63.1 45.8 68.6

99 64.1 52.0 78.0



#56

Ethyl methacrylate

Concen: 18.641 ug/l

RT: 10.877 min Scan# 1517

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21




Tgt Ion: 69 Resp: 64337


Ion Ratio Lower Upper

69 100

41 81.2 63.4 95.2

39 48.2 37.4 56.0









#57
1,3-Dichloropropane
Concen: 19.375 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 74642
Ion Ratio Lower Upper
76 100
78 32.2 25.6 38.4



#58
2-Chloroethyl Vinyl ether
Concen: 89.358 ug/l
RT: 10.165 min Scan# 1396
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion: 63 Resp: 163151
Ion Ratio Lower Upper
63 100
106 26.0 21.4 32.0









#59
2-Hexanone
Concen: 98.218 ug/l
RT: 11.200 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 280255
Ion Ratio Lower Upper
43 100
58 47.6 24.6 73.8



#60
Dibromochloromethane
Concen: 19.359 ug/l
RT: 11.365 min Scan# 1600
Delta R.T. 0.006 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion:129 Resp: 49246
Ion Ratio Lower Upper
129 100
127 78.5 39.2 117.6









#61
1,2-Dibromoethane
Concen: 19.398 ug/l
RT: 11.471 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 42620
Ion Ratio Lower Upper
107 100
109 91.5 73.7 110.5



#62
4-Bromofluorobenzene
Concen: 50.281 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion: 95 Resp: 134835
Ion Ratio Lower Upper
95 100
174 76.8 0.0 159.2
176 75.7 0.0 147.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.871 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:117 Resp: 279200

Ion Ratio Lower Upper

117 100

82 58.3 47.5 71.3

119 33.5 25.6 38.4



#64

Tetrachloroethene

Concen: 17.433 ug/l

RT: 11.106 min Scan# 1556

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Tgt Ion:164 Resp: 40131




Ion Ratio Lower Upper


164 100

166 131.6 101.4 152.0

129 91.1 75.7 113.5

131 92.4 73.8 110.6









#65
Chlorobenzene
Concen: 18.295 ug/l
RT: 11.894 min Scan# 1111
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:112 Resp: 117133
Ion Ratio Lower Upper
112 100
114 30.9 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 18.548 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion:131 Resp: 42346
Ion Ratio Lower Upper
131 100
133 96.0 47.3 142.0
119 65.0 32.5 97.4









#67
Ethyl Benzene
Concen: 17.997 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :

Tgt Ion: 91 Resp: 196333
Ion Ratio Lower Upper
91 100
106 29.8 24.0 36.0



#68
m/p-Xylenes
Concen: 37.236 ug/l
RT: 12.076 min Scan# 1721
Delta R.T. 0.006 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion:106 Resp: 151126
Ion Ratio Lower Upper
106 100
91 204.0 166.1 249.1









#69
o-Xylene
Concen: 18.454 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 70366
Ion Ratio Lower Upper
106 100
91 218.6 111.3 333.9



#70
Styrene
Concen: 18.914 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion:104 Resp: 122407
Ion Ratio Lower Upper
104 100
78 52.3 41.6 62.4
103 54.2 44.0 66.0





#71

Bromoform

Concen: 17.950 ug/l

RT: 12.582 min Scan# 1807

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:173 Resp: 32400

Ion Ratio Lower Upper

173 100

175 47.7 24.0 72.0

254 0.0 0.0 0.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.794 min Scan# 2013

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21




Tgt Ion:152 Resp: 130790


Ion Ratio Lower Upper

152 100

115 59.3 30.6 91.6

150 165.3 0.0 348.6









#73
Isopropylbenzene
Concen: 18.264 ug/l
RT: 12.700 min Scan# 1827
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 181393
Ion Ratio Lower Upper
105 100
120 25.2 12.9 38.6



#74
N-ethyl acetate
Concen: 18.206 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion: 43 Resp: 96498
Ion Ratio Lower Upper
43 100
70 37.6 29.8 44.6
55 22.9 18.7 28.1
61 21.3 17.4 26.0





#75

1,1,2-Tetrachloroethane

Concen: 18.741 ug/l

RT: 12.941 min Scan# 1868

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 83 Resp: 60714

Ion Ratio Lower Upper

83 100

131 10.2 5.0 14.9

85 65.6 31.9 95.7



#76

1,2,3-Trichloropropane

Concen: 25.950 ug/l

RT: 12.994 min Scan# 1877

Delta R.T. 0.000 min

Lab File: VN082632.D




Acq: 19 Jun 2024 11:21


Tgt Ion: 75 Resp: 83013

Ion Ratio Lower Upper

75 100

77 124.7 110.9 332.6





#77

Bromobenzene

Concen: 18.252 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:156 Resp: 46660

Ion Ratio Lower Upper

156 100

77 221.8 111.3 333.8

158 96.3 48.3 144.9



#78

n-propylbenzene

Concen: 18.303 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. 0.000 min

Lab File: VN082632.D




Acq: 19 Jun 2024 11:21


Tgt Ion: 91 Resp: 215018

Ion Ratio Lower Upper

91 100

120 21.6 10.9 32.9





#79

2-Chlorotoluene

Concen: 18.126 ug/l

RT: 13.129 min Scan# 1900

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 91 Resp: 133535

Ion Ratio Lower Upper

91 100

126 32.4 16.2 48.6



#80

1,3,5-Trimethylbenzene

Concen: 18.710 ug/l

RT: 13.176 min Scan# 1908

Delta R.T. 0.000 min

Lab File: VN082632.D




Acq: 19 Jun 2024 11:21


Tgt Ion:105 Resp: 150795

Ion Ratio Lower Upper

105 100


120 48.6 24.3 72.8








#81
trans-1,4-Dichloro-2-butene
Concen: 16.947 ug/l
RT: 12.741 min Scan# 1834
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion	Ratio	Lower	Upper
75	100		
53	126.8	88.4	132.6
89	44.6	36.3	54.5



#82
4-Chlorotoluene
Concen: 18.192 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion	Ratio	Lower	Upper
91	100		
126	32.7	16.2	48.6









#83
tert-Butylbenzene
Concen: 18.260 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:119 Resp: 125358
Ion Ratio Lower Upper
119 100
91 69.5 35.0 105.1
134 26.8 13.3 39.8



#84
1,2,4-Trimethylbenzene
Concen: 18.880 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21

Tgt Ion:105 Resp: 152648
Ion Ratio Lower Upper
105 100
120 43.5 21.9 65.8





#85

sec-Butylbenzene

Concen: 18.140 ug/l

RT: 13.618 min Scan# 1983

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:105 Resp: 176874

Ion Ratio Lower Upper

105 100

134 20.5 9.8 29.3



#86

p-Isopropyltoluene

Concen: 18.649 ug/l

RT: 13.729 min Scan# 2002

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21




Tgt Ion:119 Resp: 146428


Ion Ratio Lower Upper

119 100

134 25.0 13.0 39.0


91 24.4 12.3 36.9








#87
1,3-Dichlorobenzene
Concen: 18.013 ug/l
RT: 13.735 min Scan# 2003
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:146 Resp: 83493
Ion Ratio Lower Upper
146 100
111 41.2 20.6 61.9
148 64.7 31.8 95.3




#88
1,4-Dichlorobenzene
Concen: 17.174 ug/l
RT: 13.812 min Scan# 2016
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21



Tgt Ion:146 Resp: 83206
Ion Ratio Lower Upper
146 100
111 41.4 20.1 60.3
148 64.2 32.2 96.6





#89

n-Butylbenzene

Concen: 17.313 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 91 Resp: 120579

Ion Ratio Lower Upper

91 100

92 51.4 26.1 78.3

134 24.3 12.1 36.3



#90

Hexachloroethane

Concen: 18.624 ug/l

RT: 14.335 min Scan# 2105

Delta R.T. 0.000 min

Lab File: VN082632.D




Acq: 19 Jun 2024 11:21


Tgt Ion: 117 Resp: 31816

Ion Ratio Lower Upper

117 100


201 70.2 35.8 107.3








#91
1,2-Dichlorobenzene
Concen: 17.682 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:146 Resp: 80025
Ion Ratio Lower Upper
146 100
111 42.0 21.4 64.2
148 65.1 32.1 96.5




#92
1,2-Dibromo-3-Chloropropane
Concen: 18.882 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.000 min
Lab File: VN082632.D
Acq: 19 Jun 2024 11:21



Tgt Ion: 75 Resp: 12946
Ion Ratio Lower Upper
75 100
155 69.8 36.6 109.8
157 95.3 46.9 140.6





#93

1,2,4-Trichlorobenzene

Concen: 16.495 ug/l

RT: 15.394 min Scan# 21

Delta R.T. -0.000 min


Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :




MSVOA_N

ClientSampleId :



Tgt Ion:180 Resp: 39365

Ion	Ratio	Lower	Upper
180	100		
182	95.6	47.3	141.8
145	30.4	15.7	47.1



#94

Hexachlorobutadiene

Concen: 15.046 ug/l

RT: 15.506 min Scan# 2304




Delta R.T. 0.001 min


Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Tgt Ion:225 Resp: 17704

Ion	Ratio	Lower	Upper
225	100		
223	65.5	31.9	95.7
227	67.6	32.5	97.5





#95

Naphthalene

Concen: 17.189 ug/l

RT: 15.641 min Scan# 21

Delta R.T. 0.000 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion:128 Resp: 126206

Ion Ratio Lower Upper

128 100

127 13.2 10.6 16.0

129 10.4 8.6 12.8



#96

1,2,3-Trichlorobenzene

Concen: 17.434 ug/l

RT: 15.847 min Scan# 2362

Delta R.T. 0.006 min

Lab File: VN082632.D

Acq: 19 Jun 2024 11:21

Tgt Ion:180 Resp: 42401

Ion Ratio Lower Upper

180 100

182 93.2 48.9 146.8

145 32.5 16.8 50.4

