

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
 Data File : VN082636.D
 Acq On : 19 Jun 2024 13:08
 Operator : JC\MD
 Sample : PB161542ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161542ZHE#02

Quant Time: Jun 20 02:59:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	132244	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	252073	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	259069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103936	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	121491	57.288	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.580%
35) Dibromofluoromethane	8.171	113	88870	51.970	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.940%
50) Toluene-d8	10.571	98	339638	52.521	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.040%
62) 4-Bromofluorobenzene	12.853	95	110784	49.690	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	99.380%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	58860	61.028	ug/l	98
20) Methylene Chloride	5.277	84	11657	6.875	ug/l	96
25) 2-Butanone	7.494	43	17352	12.658	ug/l	97
43) Isopropyl Acetate	8.694	43	234248	38.493	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
Data File : VN082636.D
Acq On : 19 Jun 2024 13:08
Operator : JC\MD
Sample : PB161542ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161542ZHE#02

Quant Time: Jun 20 02:59:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

