

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
 Data File : VN082641.D
 Acq On : 19 Jun 2024 15:11
 Operator : JC\MD
 Sample : PB161542ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161542ZHE#07

Quant Time: Jun 20 03:02:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	131007	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	250894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	260959	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	106622	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	120358	57.289	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.580%
35) Dibromofluoromethane	8.171	113	89203	52.409	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.820%
50) Toluene-d8	10.571	98	335793	52.170	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.340%
62) 4-Bromofluorobenzene	12.853	95	110315	49.712	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	99.420%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	23142	24.221	ug/l	93
20) Methylene Chloride	5.283	84	17747	10.565	ug/l	92
25) 2-Butanone	7.489	43	14016	10.321	ug/l	93
43) Isopropyl Acetate	8.694	43	268631	44.350	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
Data File : VN082641.D
Acq On : 19 Jun 2024 15:11
Operator : JC\MD
Sample : PB161542ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161542ZHE#07

Quant Time: Jun 20 03:02:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

