

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
 Data File : VN082642.D
 Acq On : 19 Jun 2024 15:36
 Operator : JC\MD
 Sample : PB161542ZHE#08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB161542ZHE#08

Quant Time: Jun 20 03:02:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	129253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	249773	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	256568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	102599	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	120481	58.126	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 116.260%			
35) Dibromofluoromethane	8.171	113	89009	52.530	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.060%			
50) Toluene-d8	10.571	98	333149	51.992	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.980%			
62) 4-Bromofluorobenzene	12.853	95	108596	49.157	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 98.320%			
Target Compounds					Qvalue	
16) Acetone	4.441	43	23329	24.748	ug/l	92
20) Methylene Chloride	5.283	84	17402	10.500	ug/l	92
25) 2-Butanone	7.488	43	14271	10.651	ug/l	93
43) Isopropyl Acetate	8.694	43	275812	45.740	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
Data File : VN082642.D
Acq On : 19 Jun 2024 15:36
Operator : JC\MD
Sample : PB161542ZHE#08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB161542ZHE#08

Quant Time: Jun 20 03:02:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

