

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
 Data File : VN082646.D
 Acq On : 19 Jun 2024 17:14
 Operator : JC\MD
 Sample : P2915-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-NORTH-1

Quant Time: Jun 20 03:03:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	126453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	247773	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	255358	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101359	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	120396	59.371	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.740%
35) Dibromofluoromethane	8.171	113	87172	51.861	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.720%
50) Toluene-d8	10.571	98	332950	52.380	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.760%
62) 4-Bromofluorobenzene	12.853	95	106003	48.371	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	96.740%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	20727	22.475	ug/l	90
18) Methyl Acetate	5.030	43	3885	1.722	ug/l	97
20) Methylene Chloride	5.283	84	10152	6.261	ug/l #	89
25) 2-Butanone	7.489	43	11863	9.050	ug/l	99
43) Isopropyl Acetate	8.694	43	219377	36.675	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
Data File : VN082646.D
Acq On : 19 Jun 2024 17:14
Operator : JC\MD
Sample : P2915-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-NORTH-1

Quant Time: Jun 20 03:03:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

