

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
 Data File : VN082647.D
 Acq On : 19 Jun 2024 17:39
 Operator : JC\MD
 Sample : P2915-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-SOUTH-1

Quant Time: Jun 20 03:04:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	126064	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	246119	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	255602	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	100264	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	118275	58.505	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.020%
35) Dibromofluoromethane	8.171	113	88051	52.736	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.480%
50) Toluene-d8	10.571	98	330315	52.315	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.620%
62) 4-Bromofluorobenzene	12.853	95	105811	48.607	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	97.220%
Target Compounds						
						Qvalue
16) Acetone	4.442	43	21590	23.483	ug/l	92
20) Methylene Chloride	5.271	84	10489	6.489	ug/l #	84
25) 2-Butanone	7.494	43	12745	9.753	ug/l #	89
43) Isopropyl Acetate	8.694	43	197622	33.260	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
Data File : VN082647.D
Acq On : 19 Jun 2024 17:39
Operator : JC\MD
Sample : P2915-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-SOUTH-1

Quant Time: Jun 20 03:04:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

