

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
 Data File : VN082648.D
 Acq On : 19 Jun 2024 18:04
 Operator : JC\MD
 Sample : P2930-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WCS-TPE

Quant Time: Jun 20 03:04:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	126132	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	241166	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	253466	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	99846	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	117514	58.098	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.200%
35) Dibromofluoromethane	8.171	113	86496	52.869	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.740%
50) Toluene-d8	10.571	98	329923	53.326	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.660%
62) 4-Bromofluorobenzene	12.853	95	106620	49.985	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	99.980%
Target Compounds						
						Qvalue
16) Acetone	4.442	43	28971	31.494	ug/l	98
20) Methylene Chloride	5.283	84	17638	10.906	ug/l #	79
25) 2-Butanone	7.494	43	15142	11.581	ug/l	100
43) Isopropyl Acetate	8.694	43	226964	38.983	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
Data File : VN082648.D
Acq On : 19 Jun 2024 18:04
Operator : JC\MD
Sample : P2930-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WCS-TPE

Quant Time: Jun 20 03:04:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

