

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
 Data File : VN082650.D
 Acq On : 19 Jun 2024 18:53
 Operator : JC\MD
 Sample : P2955-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Jun 20 03:05:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	125745	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	245897	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	255114	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101121	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	118786	58.907	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.820%
35) Dibromofluoromethane	8.171	113	87426	52.409	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.820%
50) Toluene-d8	10.571	98	329854	52.289	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.580%
62) 4-Bromofluorobenzene	12.853	95	105937	48.709	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	97.420%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	32086	34.987	ug/l #	89
18) Methyl Acetate	5.036	43	3380	1.507	ug/l #	70
20) Methylene Chloride	5.277	84	5304	3.290	ug/l #	86
43) Isopropyl Acetate	8.694	43	227540	38.330	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061924\
Data File : VN082650.D
Acq On : 19 Jun 2024 18:53
Operator : JC\MD
Sample : P2955-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Jun 20 03:05:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

