

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061925\
 Data File : VN087103.D
 Acq On : 19 Jun 2025 10:47
 Operator : JC\MD
 Sample : Q2329-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB01-20250613

Quant Time: Jun 20 01:21:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	167515	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	339842	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	308249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	141080	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	127623	56.902	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.800%
35) Dibromofluoromethane	8.177	113	107337	53.296	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.600%
50) Toluene-d8	10.565	98	423645	53.136	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.280%
62) 4-Bromofluorobenzene	12.847	95	144889	48.914	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.820%
Target Compounds						
						Qvalue
16) Acetone	4.459	43	13800	11.603	ug/l	96
20) Methylene Chloride	5.306	84	3318	1.490	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061925\
Data File : VN087103.D
Acq On : 19 Jun 2025 10:47
Operator : JC\MD
Sample : Q2329-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB01-20250613

Quant Time: Jun 20 01:21:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

