

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061925\
 Data File : VN087110.D
 Acq On : 19 Jun 2025 14:42
 Operator : JC\MD
 Sample : Q2333-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-13

Quant Time: Jun 20 01:25:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	200534	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	398084	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	356355	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	169301	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	136952	51.008	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.020%		
35) Dibromofluoromethane	8.177	113	121246	51.394	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.780%		
50) Toluene-d8	10.565	98	483778	51.801	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.600%		
62) 4-Bromofluorobenzene	12.847	95	171089	49.308	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 98.620%		
Target Compounds						
						Qvalue
16) Acetone	4.453	43	8627	6.059	ug/l	# 81
40) Benzene	8.618	78	9222	0.802	ug/l	97
67) Ethyl Benzene	11.965	91	22730	1.679	ug/l	97
68) m/p-Xylenes	12.071	106	7635	1.473	ug/l	97
69) o-Xylene	12.400	106	6022	1.213	ug/l	94
84) 1,2,4-Trimethylbenzene	13.482	105	7202	0.705	ug/l	96
95) Naphthalene	15.641	128	4261	0.335	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN061925\
Data File : VN087110.D
Acq On : 19 Jun 2025 14:42
Operator : JC\MD
Sample : Q2333-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-13

Quant Time: Jun 20 01:25:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

