

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN062019\
 Data File : VN056389.D
 Acq On : 20 Jun 2019 9:58
 Operator : JC/SP
 Sample : VN0620MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0620MBL01

Quant Time: Jun 21 04:39:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	356645	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	595603	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	561531	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	158690	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	197754	50.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.06%	
35) Dibromofluoromethane	7.59	113	177996	47.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.82%	
50) Toluene-d8	10.09	98	725246	50.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.32%	
62) 4-Bromofluorobenzene	12.40	95	225025	46.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.38%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN062019\
Data File : VN056389.D
Acq On : 20 Jun 2019 9:58
Operator : JC/SP
Sample : VN0620MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0620MBL01

Quant Time: Jun 21 04:39:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

