

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062019\
 Data File : VN056393.D
 Acq On : 20 Jun 2019 11:32
 Operator : JC/SP
 Sample : K3454-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 30271

Quant Time: Jun 21 04:48:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	355917	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	601585	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	599270	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	224133	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	202178	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
35) Dibromofluoromethane	7.59	113	178167	47.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.94%	
50) Toluene-d8	10.09	98	743253	50.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.80%	
62) 4-Bromofluorobenzene	12.40	95	257980	53.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.00%	
Target Compounds						
16) Acetone	3.82	43	24671	12.483	ug/l	97
17) Carbon Disulfide	4.05	76	7452	1.042	ug/l #	89
18) Methyl Acetate	4.33	43	7970	1.212	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062019\
Data File : VN056393.D
Acq On : 20 Jun 2019 11:32
Operator : JC/SP
Sample : K3454-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
30271

Quant Time: Jun 21 04:48:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

