

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062019\
 Data File : VN056398.D
 Acq On : 20 Jun 2019 13:29
 Operator : JC/SP
 Sample : K3455-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5-DUCT-BANK

Quant Time: Jun 21 05:03:09 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	380030	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	634488	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	637705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	196209	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	212234	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
35) Dibromofluoromethane	7.59	113	187722	47.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.86%	
50) Toluene-d8	10.09	98	797390	51.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.54%	
62) 4-Bromofluorobenzene	12.40	95	262008	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
Target Compounds						
20) Methylene Chloride	4.55	84	7107613	1818.177	ug/l	99
43) Isopropyl Acetate	8.17	43	14892	1.760	ug/l #	88
95) Naphthalene	15.13	128	3512	5.595	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062019\
Data File : VN056398.D
Acq On : 20 Jun 2019 13:29
Operator : JC/SP
Sample : K3455-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5-DUCT-BANK

Quant Time: Jun 21 05:03:09 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

