

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
 Data File : VN072947.D
 Acq On : 20 Jun 2022 09:03
 Operator : JC\MD
 Sample : VN0620WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0620WBL01

Quant Time: Jun 21 05:53:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	357188	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	565364	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	508715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	261403	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.363	65	203207	42.080	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.160%
35) Dibromofluoromethane	7.951	113	179909	51.484	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.960%
50) Toluene-d8	10.375	98	578177	44.465	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.940%
62) 4-Bromofluorobenzene	12.663	95	261167	55.856	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
Data File : VN072947.D
Acq On : 20 Jun 2022 09:03
Operator : JC\MD
Sample : VN0620WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0620WBL01

Quant Time: Jun 21 05:53:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

