

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
 Data File : VN072948.D
 Acq On : 20 Jun 2022 09:40
 Operator : JC\MD
 Sample : VN0620WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0620WBL02

Quant Time: Jun 21 05:53:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	349671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	552859	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.674	117	497680	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.604	152	248498	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	196996	41.671	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.340%
35) Dibromofluoromethane	7.951	113	173081	50.650	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.300%
50) Toluene-d8	10.375	98	567245	44.611	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.220%
62) 4-Bromofluorobenzene	12.663	95	247375	54.103	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
Data File : VN072948.D
Acq On : 20 Jun 2022 09:40
Operator : JC\MD
Sample : VN0620WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0620WBL02

Quant Time: Jun 21 05:53:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

