

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
 Data File : VN072952.D
 Acq On : 20 Jun 2022 11:26
 Operator : JC\MD
 Sample : N3367-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MJS-MW-01

Quant Time: Jun 21 05:54:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	355180	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	563069	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	502338	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	245642	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	204421	42.570	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.140%
35) Dibromofluoromethane	7.951	113	182136	52.334	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.660%
50) Toluene-d8	10.374	98	577044	44.559	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.120%
62) 4-Bromofluorobenzene	12.668	95	250341	53.759	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.520%
Target Compounds						
					Qvalue	
30) Chloroform	7.745	83	361492	50.287	ug/l	97
47) Bromodichloromethane	9.692	83	19499	3.836	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
Data File : VN072952.D
Acq On : 20 Jun 2022 11:26
Operator : JC\MD
Sample : N3367-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MJS-MW-01

Quant Time: Jun 21 05:54:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

