

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
 Data File : VN072960.D
 Acq On : 20 Jun 2022 14:40
 Operator : JC\MD
 Sample : N3357-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-7

Quant Time: Jun 21 05:55:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	332757	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	526763	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	479005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	231935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	195259	43.403	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.800%
35) Dibromofluoromethane	7.951	113	166358	51.095	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.180%
50) Toluene-d8	10.375	98	545727	45.045	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.080%
62) 4-Bromofluorobenzene	12.669	95	237396	54.493	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.980%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	6737	2.724	ug/l #	73
67) Ethyl Benzene	11.775	91	6888	0.418	ug/l	100
73) Isopropylbenzene	12.516	105	4752	0.265	ug/l	99
78) n-propylbenzene	12.857	91	7083	0.359	ug/l	93
95) Naphthalene	15.427	128	7079	0.395	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
Data File : VN072960.D
Acq On : 20 Jun 2022 14:40
Operator : JC\MD
Sample : N3357-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-7

Quant Time: Jun 21 05:55:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

