

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
 Data File : VN072963.D
 Acq On : 20 Jun 2022 15:52
 Operator : JC\MD
 Sample : N3357-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-2022-06-14

Quant Time: Jun 21 05:55:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	362966	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	565043	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	498833	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.609	152	245990	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	208066	42.400	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	84.800%	
35) Dibromofluoromethane	7.951	113	179041	51.265	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.520%	
50) Toluene-d8	10.374	98	575882	44.314	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	88.620%	
62) 4-Bromofluorobenzene	12.662	95	248133	53.099	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	106.200%	
Target Compounds						
					Qvalue	
16) Acetone	4.239	43	13324	4.939	ug/l	93
67) Ethyl Benzene	11.780	91	42850	2.499	ug/l	98
68) m/p-Xylenes	11.886	106	22021	3.274	ug/l	99
78) n-propylbenzene	12.851	91	15086	0.721	ug/l	93
80) 1,3,5-Trimethylbenzene	12.998	105	9045	0.579	ug/l	87
84) 1,2,4-Trimethylbenzene	13.304	105	66630	4.279	ug/l	96
95) Naphthalene	15.427	128	227814	11.988	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 21 05:55:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

