

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
 Data File : VN072966.D
 Acq On : 20 Jun 2022 17:05
 Operator : JC\MD
 Sample : N3370-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-07

Quant Time: Jun 21 05:56:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	341622	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	545811	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	494634	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.609	152	249305	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	199982	43.299	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.600%		
35) Dibromofluoromethane	7.951	113	168630	49.985	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.980%		
50) Toluene-d8	10.374	98	563922	44.922	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	89.840%		
62) 4-Bromofluorobenzene	12.668	95	247569	54.845	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	109.680%		
Target Compounds						Qvalue
11) Tert butyl alcohol	5.310	59	13898	11.649	ug/l	99
19) Methyl tert-butyl Ether	5.539	73	51409	4.185	ug/l	97
22) Diisopropyl ether	6.410	45	8690	0.667	ug/l #	94
27) cis-1,2-Dichloroethene	7.251	96	4329	1.018	ug/l	86
67) Ethyl Benzene	11.780	91	5999	0.353	ug/l	90
68) m/p-Xylenes	11.886	106	3278	0.491	ug/l	99
84) 1,2,4-Trimethylbenzene	13.304	105	9896	0.627	ug/l	96
95) Naphthalene	15.433	128	20924	1.086	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
Data File : VN072966.D
Acq On : 20 Jun 2022 17:05
Operator : JC\MD
Sample : N3370-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-07

Quant Time: Jun 21 05:56:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

