

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
 Data File : VN072967.D
 Acq On : 20 Jun 2022 17:30
 Operator : JC\MD
 Sample : N3370-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TOLL-MW-08

Quant Time: Jun 21 05:56:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	339442	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.899	114	544429	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.681	117	497045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	252691	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	203477	44.339	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	88.680%		
35) Dibromofluoromethane	7.952	113	171872	51.075	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	102.160%		
50) Toluene-d8	10.375	98	559487	44.682	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	89.360%		
62) 4-Bromofluorobenzene	12.669	95	252402	56.058	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	112.120%		
Target Compounds						Qvalue
19) Methyl tert-butyl Ether	5.552	73	4739	0.388	ug/l	99
67) Ethyl Benzene	11.781	91	4772	0.279	ug/l	91
68) m/p-Xylenes	11.881	106	2277	0.340	ug/l	97
84) 1,2,4-Trimethylbenzene	13.298	105	5977	0.374	ug/l	100
95) Naphthalene	15.433	128	14000	0.717	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
Data File : VN072967.D
Acq On : 20 Jun 2022 17:30
Operator : JC\MD
Sample : N3370-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TOLL-MW-08

Quant Time: Jun 21 05:56:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

