

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
 Data File : VN072959.D
 Acq On : 20 Jun 2022 14:16
 Operator : JC\MD
 Sample : N3357-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-5

Quant Time: Jun 21 05:55:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	326389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	513158	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	462890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	223739	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	188472	42.711	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.420%
35) Dibromofluoromethane	7.951	113	164164	51.758	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.520%
50) Toluene-d8	10.375	98	536403	45.449	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.900%
62) 4-Bromofluorobenzene	12.663	95	230194	54.241	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.480%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062022\
Data File : VN072959.D
Acq On : 20 Jun 2022 14:16
Operator : JC\MD
Sample : N3357-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-5

Quant Time: Jun 21 05:55:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.016 min Scan# 1042
Delta R.T. -0.000 min
Lab File: VN072959.D
Acq: 20 Jun 2022 14:16




Instrument :
MSVOA_N
ClientSampleId :
MW-5


Tgt Ion:168 Resp: 326389
Ion Ratio Lower Upper
168 100
99 57.3 42.3 63.5



#33
1,2-Dichloroethane-d4
Concen: 42.711 ug/l
RT: 8.369 min Scan# 1102
Delta R.T. 0.000 min
Lab File: VN072959.D
Acq: 20 Jun 2022 14:16

Tgt Ion: 65 Resp: 188472
Ion Ratio Lower Upper
65 100
67 51.6 0.0 104.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.904 min Scan# 1192

Delta R.T. 0.006 min

Lab File: VN072959.D

Acq: 20 Jun 2022 14:16

Instrument :

MSVOA_N

ClientSampleId :

MW-5





Tgt Ion:114 Resp: 513158

Ion Ratio Lower Upper

114 100

63 20.4 0.0 40.0

88 15.7 0.0 32.6



#35

Dibromofluoromethane

Concen: 51.758 ug/l

RT: 7.951 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VN072959.D

Acq: 20 Jun 2022 14:16




Tgt Ion:113 Resp: 164164


Ion Ratio Lower Upper

113 100

111 102.8 81.4 122.0

192 18.8 17.0 25.6









#50
Toluene-d8
Concen: 45.449 ug/l
RT: 10.375 min Scan# 1443
Delta R.T. 0.000 min
Lab File: VN072959.D
Acq: 20 Jun 2022 14:16




Instrument :
MSVOA_N
ClientSampleId :
MW-5


Tgt Ion: 98 Resp: 536403
Ion Ratio Lower Upper
98 100
100 64.8 52.6 78.8



#62
4-Bromofluorobenzene
Concen: 54.241 ug/l
RT: 12.663 min Scan# 1832
Delta R.T. -0.006 min
Lab File: VN072959.D
Acq: 20 Jun 2022 14:16

Tgt Ion: 95 Resp: 230194
Ion Ratio Lower Upper
95 100
174 76.0 0.0 167.6
176 72.9 0.0 163.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.680 min Scan# 1166
Delta R.T. 0.000 min
Lab File: VN072959.D
Acq: 20 Jun 2022 14:16

Instrument :
MSVOA_N
ClientSampleId :
MW-5

Tgt Ion:117 Resp: 462890
Ion Ratio Lower Upper
117 100
82 55.2 44.2 66.2
119 32.4 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.610 min Scan# 1993
Delta R.T. 0.000 min
Lab File: VN072959.D
Acq: 20 Jun 2022 14:16

Tgt Ion:152 Resp: 223739
Ion Ratio Lower Upper
152 100
115 59.6 28.9 86.7
150 156.9 0.0 356.2

