

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062023\
 Data File : VN078298.D
 Acq On : 20 Jun 2023 10:26
 Operator : JC\MD
 Sample : VN0620MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0620MBL01

Quant Time: Jun 21 05:42:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	427933	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	786797	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	682141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	202965	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	300570	52.679	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.360%
35) Dibromofluoromethane	8.177	113	275325	55.344	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.680%
50) Toluene-d8	10.570	98	911712	47.538	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.080%
62) 4-Bromofluorobenzene	12.853	95	278779	46.759	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062023\
Data File : VN078298.D
Acq On : 20 Jun 2023 10:26
Operator : JC\MD
Sample : VN0620MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0620MBL01

Quant Time: Jun 21 05:42:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

