

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062023\
 Data File : VN078305.D
 Acq On : 20 Jun 2023 13:24
 Operator : JC\MD
 Sample : 03321-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-24

Quant Time: Jun 21 05:45:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	668030	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	867235	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	805653	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	258381	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	330040	37.054	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	74.100%
35) Dibromofluoromethane	8.171	113	404507	73.769	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	147.540%#
50) Toluene-d8	10.571	98	1043093	49.344	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.680%
62) 4-Bromofluorobenzene	12.853	95	335579	51.065	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.120%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	69122	26.962	ug/l	91
20) Methylene Chloride	5.283	84	22418	2.646	ug/l #	87
37) Ethyl Acetate	7.571	43	74271	12.088	ug/l	96
43) Isopropyl Acetate	8.694	43	282653	24.054	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062023\
Data File : VN078305.D
Acq On : 20 Jun 2023 13:24
Operator : JC\MD
Sample : 03321-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-24

Quant Time: Jun 21 05:45:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

