

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062023\
 Data File : VN078311.D
 Acq On : 20 Jun 2023 15:49
 Operator : JC\MD
 Sample : 03321-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-22

Quant Time: Jun 21 05:48:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	429029	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	820500	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	761417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	234056	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	313424	54.792	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.580%		
35) Dibromofluoromethane	8.177	113	285091	54.953	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 109.900%		
50) Toluene-d8	10.571	98	969278	48.464	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 96.920%		
62) 4-Bromofluorobenzene	12.853	95	309463	49.773	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 99.540%		
Target Compounds						Qvalue
16) Acetone	4.447	43	35559	21.597	ug/l	100
20) Methylene Chloride	5.289	84	15390	2.829	ug/l #	84
37) Ethyl Acetate	7.571	43	60226	10.360	ug/l	96
43) Isopropyl Acetate	8.700	43	372084	33.468	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062023\
Data File : VN078311.D
Acq On : 20 Jun 2023 15:49
Operator : JC\MD
Sample : 03321-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-22

Quant Time: Jun 21 05:48:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

