

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062023\
 Data File : VN078312.D
 Acq On : 20 Jun 2023 16:13
 Operator : JC\MD
 Sample : 03321-05RE
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-24

Quant Time: Jun 21 05:48:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	421045	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	792964	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	726505	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	220459	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	301660	53.735	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.480%			
35) Dibromofluoromethane	8.177	113	277273	55.302	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 110.600%			
50) Toluene-d8	10.576	98	963672	49.857	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.720%			
62) 4-Bromofluorobenzene	12.853	95	294828	49.066	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 98.140%			
Target Compounds						
16) Acetone	4.447	43	52871	32.720	ug/l	90
20) Methylene Chloride	5.295	84	16643	3.117	ug/l	96
25) 2-Butanone	7.500	43	13127	5.283	ug/l	88
37) Ethyl Acetate	7.577	43	58082	10.338	ug/l #	96
43) Isopropyl Acetate	8.700	43	386167	35.941	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062023\
Data File : VN078312.D
Acq On : 20 Jun 2023 16:13
Operator : JC\MD
Sample : 03321-05RE
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-24

Quant Time: Jun 21 05:48:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

