

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062024\
 Data File : VN082662.D
 Acq On : 20 Jun 2024 12:45
 Operator : JC\MD
 Sample : P2981-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 240619057-03-TRIP-BLANK

Quant Time: Jun 21 02:04:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	136297	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	248487	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	245852	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	102270	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	121866	55.756	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 111.520%	
35) Dibromofluoromethane	8.171	113	90778	53.851	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.700%	
50) Toluene-d8	10.571	98	325038	50.988	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.980%	
62) 4-Bromofluorobenzene	12.853	95	103141	46.929	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 93.860%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062024\
Data File : VN082662.D
Acq On : 20 Jun 2024 12:45
Operator : JC\MD
Sample : P2981-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
240619057-03-TRIP-BLANK

Quant Time: Jun 21 02:04:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

