

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062024\
 Data File : VN082663.D
 Acq On : 20 Jun 2024 13:09
 Operator : JC\MD
 Sample : P2978-01 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWBR-05-115-142-061924

Quant Time: Jun 21 02:05:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	134426	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	247482	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	242547	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	103079	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	122903	57.013	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.020%
35) Dibromofluoromethane	8.165	113	91910	54.744	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.480%
50) Toluene-d8	10.571	98	322258	50.758	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.520%
62) 4-Bromofluorobenzene	12.853	95	105779	48.325	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	96.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062024\
Data File : VN082663.D
Acq On : 20 Jun 2024 13:09
Operator : JC\MD
Sample : P2978-01 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWBR-05-115-142-061924

Quant Time: Jun 21 02:05:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

