

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062024\
 Data File : VN082668.D
 Acq On : 20 Jun 2024 15:12
 Operator : JC\MD
 Sample : P2993-03 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 528A

Quant Time: Jun 21 02:07:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	122491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	226417	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	221741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	86473	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	114490	58.285	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.580%
35) Dibromofluoromethane	8.171	113	84704	55.146	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.300%
50) Toluene-d8	10.571	98	299433	51.550	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.100%
62) 4-Bromofluorobenzene	12.853	95	92170	46.025	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	92.060%
Target Compounds						
						Qvalue
16) Acetone	4.442	43	7780	8.709	ug/l	94
52) Toluene	10.629	92	6657	1.548	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062024\
Data File : VN082668.D
Acq On : 20 Jun 2024 15:12
Operator : JC\MD
Sample : P2993-03 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
528A

Quant Time: Jun 21 02:07:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

