

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062024\
 Data File : VN082656.D
 Acq On : 20 Jun 2024 10:00
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 21 02:00:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.698	0.663	5.0	85	0.00
3 P	Chloromethane	0.837	0.752	10.2	89	0.00
4 C	Vinyl Chloride	0.797	0.734	7.9#	88	0.00
5 T	Bromomethane	0.513	0.454	11.5	89	0.00
6 T	Chloroethane	0.568	0.539	5.1	96	0.00
7 T	Trichlorofluoromethane	1.039	0.991	4.6	92	0.00
8 T	Diethyl Ether	0.354	0.339	4.2	89	-0.01
9 T	1,1,2-Trichlorotrifluoroeth	0.582	0.574	1.4	94	0.00
10 T	Methyl Iodide	0.555	0.616	-11.0	97	0.00
11 T	Tert butyl alcohol	0.131	0.130	0.8	94	0.00
12 CM	1,1-Dichloroethene	0.557	0.542	2.7#	94	0.00
13 T	Acrolein	0.134	0.117	12.7	85	0.00
14 T	Allyl chloride	1.089	0.993	8.8	88	0.00
15 T	Acrylonitrile	0.346	0.346	0.0	97	0.00
16 T	Acetone	0.365	0.346	5.2	92	0.00
17 T	Carbon Disulfide	1.768	1.574	11.0	88	0.00
18 T	Methyl Acetate	0.892	0.873	2.1	99	0.00
19 T	Methyl tert-butyl Ether	1.906	1.964	-3.0	92	0.00
20 T	Methylene Chloride	0.641	0.641	0.0	97	0.00
21 T	trans-1,2-Dichloroethene	0.592	0.586	1.0	94	0.00
22 T	Diisopropyl ether	2.283	2.378	-4.2	94	0.00
23 T	Vinyl Acetate	1.811	1.890	-4.4	93	0.00
24 P	1,1-Dichloroethane	1.237	1.253	-1.3	97	0.00
25 T	2-Butanone	0.518	0.525	-1.4	95	0.00
26 T	2,2-Dichloropropane	1.073	1.067	0.6	92	0.00
27 T	cis-1,2-Dichloroethene	0.713	0.700	1.8	93	0.00
28 T	Bromochloromethane	0.564	0.566	-0.4	101	0.00
29 T	Tetrahydrofuran	0.317	0.343	-8.2	98	0.00
30 C	Chloroform	1.207	1.239	-2.7#	98	0.00
31 T	Cyclohexane	1.157	1.049	9.3	88	0.00
32 T	1,1,1-Trichloroethane	1.094	1.090	0.4	95	0.00
33 S	1,2-Dichloroethane-d4	0.802	0.830	-3.5	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.339	0.360	-6.2	102	0.00
36 T	1,1-Dichloropropene	0.507	0.536	-5.7	95	0.00
37 T	Ethyl Acetate	0.646	0.641	0.8	98	0.00
38 T	Carbon Tetrachloride	0.571	0.593	-3.9	97	0.00
39 T	Methylcyclohexane	0.534	0.561	-5.1	90	0.00
40 TM	Benzene	1.567	1.638	-4.5	96	0.00
41 T	Methacrylonitrile	0.330	0.344	-4.2	96	0.00
42 TM	1,2-Dichloroethane	0.596	0.623	-4.5	98	0.00
43 T	Isopropyl Acetate	1.207	1.164	3.6	99	0.00
44 TM	Trichloroethene	0.364	0.367	-0.8	94	0.00
45 C	1,2-Dichloropropane	0.403	0.434	-7.7#	99	0.00
46 T	Dibromomethane	0.273	0.291	-6.6	98	0.00
47 T	Bromodichloromethane	0.583	0.616	-5.7	98	0.00
48 T	Methyl methacrylate	0.498	0.518	-4.0	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062024\
 Data File : VN082656.D
 Acq On : 20 Jun 2024 10:00
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 21 02:00:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	99	0.00
50 S	Toluene-d8	1.283	1.410	-9.9	100	0.00
51 T	4-Methyl-2-Pentanone	0.619	0.681	-10.0	99	0.00
52 CM	Toluene	0.949	1.043	-9.9#	97	0.00
53 T	t-1,3-Dichloropropene	0.596	0.638	-7.0	93	0.00
54 T	cis-1,3-Dichloropropene	0.616	0.670	-8.8	94	0.00
55 T	1,1,2-Trichloroethane	0.344	0.381	-10.8	99	0.00
56 T	Ethyl methacrylate	0.569	0.642	-12.8	96	0.00
57 T	1,3-Dichloropropane	0.635	0.689	-8.5	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.301	0.333	-10.6	99	0.00
59 T	2-Hexanone	0.471	0.519	-10.2	96	0.00
60 T	Dibromochloromethane	0.420	0.467	-11.2	99	0.00
61 T	1,2-Dibromoethane	0.362	0.381	-5.2	96	0.00
62 S	4-Bromofluorobenzene	0.442	0.494	-11.8	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.412	0.381	7.5	94	0.00
65 PM	Chlorobenzene	1.147	1.171	-2.1	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.409	0.425	-3.9	98	0.00
67 C	Ethyl Benzene	1.954	2.103	-7.6#	96	0.00
68 T	m/p-Xylenes	0.727	0.803	-10.5	96	0.00
69 T	o-Xylene	0.683	0.749	-9.7	95	0.00
70 T	Styrene	1.159	1.345	-16.0	97	0.00
71 P	Bromoform	0.323	0.345	-6.8	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.797	3.781	0.4	94	0.00
74 T	N-amyl acetate	2.026	1.992	1.7	94	0.00
75 P	1,1,2,2-Tetrachloroethane	1.238	1.178	4.8	102	0.00
76 T	1,2,3-Trichloropropane	1.223	1.118	8.6	98	0.00
77 T	Bromobenzene	0.977	0.946	3.2	98	0.00
78 T	n-propylbenzene	4.491	4.694	-4.5	96	0.00
79 T	2-Chlorotoluene	2.816	2.804	0.4	96	0.00
80 T	1,3,5-Trimethylbenzene	3.081	3.244	-5.3	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.511	0.533	-4.3	107	0.00
82 T	4-Chlorotoluene	2.813	2.818	-0.2	95	0.00
83 T	tert-Butylbenzene	2.625	2.706	-3.1	95	0.00
84 T	1,2,4-Trimethylbenzene	3.091	3.321	-7.4	96	0.00
85 T	sec-Butylbenzene	3.727	3.859	-3.5	94	0.00
86 T	p-Isopropyltoluene	3.002	3.246	-8.1	94	0.00
87 T	1,3-Dichlorobenzene	1.772	1.764	0.5	97	0.00
88 T	1,4-Dichlorobenzene	1.852	1.765	4.7	97	0.00
89 T	n-Butylbenzene	2.663	2.754	-3.4	91	0.00
90 T	Hexachloroethane	0.653	0.633	3.1	97	0.00
91 T	1,2-Dichlorobenzene	1.730	1.689	2.4	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.262	0.248	5.3	94	0.00
93 T	1,2,4-Trichlorobenzene	0.912	0.867	4.9	91	0.00
94 T	Hexachlorobutadiene	0.450	0.381	15.3	90	0.00
95 T	Naphthalene	2.807	2.864	-2.0	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062024\
Data File : VN082656.D
Acq On : 20 Jun 2024 10:00
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 21 02:00:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.930	0.893	4.0	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6