

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062024\
 Data File : VN082658.D
 Acq On : 20 Jun 2024 10:58
 Operator : JC\MD
 Sample : VN0620WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0620WBL01

Quant Time: Jun 21 02:02:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	139058	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	256392	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	249585	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	104962	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	124781	55.956	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.920%
35) Dibromofluoromethane	8.165	113	93286	53.633	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.260%
50) Toluene-d8	10.571	98	329376	50.076	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.160%
62) 4-Bromofluorobenzene	12.853	95	104959	46.284	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	92.560%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062024\
Data File : VN082658.D
Acq On : 20 Jun 2024 10:58
Operator : JC\MD
Sample : VN0620WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0620WBL01

Quant Time: Jun 21 02:02:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1067
Delta R.T. 0.001 min
Lab File: VN082658.D
Acq: 20 Jun 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VN0620WBL01


Tgt Ion:168 Resp: 139058
Ion Ratio Lower Upper
168 100
99 57.0 48.2 72.4



#33
1,2-Dichloroethane-d4
Concen: 55.956 ug/l
RT: 8.583 min Scan# 1127
Delta R.T. 0.000 min
Lab File: VN082658.D
Acq: 20 Jun 2024 10:58

Tgt Ion: 65 Resp: 124781
Ion Ratio Lower Upper
65 100
67 50.9 0.0 104.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN082658.D

Acq: 20 Jun 2024 10:58

Instrument :

MSVOA_N

ClientSampleId :

VN0620WBL01





Tgt Ion:114 Resp: 256392

Ion Ratio Lower Upper

114 100

63 22.2 0.0 44.6

88 17.2 0.0 31.4



#35

Dibromofluoromethane

Concen: 53.633 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN082658.D

Acq: 20 Jun 2024 10:58




Tgt Ion:113 Resp: 93286


Ion Ratio Lower Upper

113 100

111 103.3 82.4 123.6

192 18.8 14.9 22.3









#50
Toluene-d8
Concen: 50.076 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VN082658.D
Acq: 20 Jun 2024 10:58




Instrument :
MSVOA_N
ClientSampleId :
VN0620WBL01


Tgt Ion: 98 Resp: 329376
Ion Ratio Lower Upper
98 100
100 65.3 51.5 77.3



#62
4-Bromofluorobenzene
Concen: 46.284 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.006 min
Lab File: VN082658.D
Acq: 20 Jun 2024 10:58

Tgt Ion: 95 Resp: 104959
Ion Ratio Lower Upper
95 100
174 79.8 0.0 159.2
176 77.6 0.0 147.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN082658.D
Acq: 20 Jun 2024 10:58

Instrument :
MSVOA_N
ClientSampleId :
VN0620WBL01

Tgt Ion:117 Resp: 249585
Ion Ratio Lower Upper
117 100
82 59.5 47.5 71.3
119 34.0 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN082658.D
Acq: 20 Jun 2024 10:58

Tgt Ion:152 Resp: 104962
Ion Ratio Lower Upper
152 100
115 60.0 30.6 91.6
150 157.7 0.0 348.6

