

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062025\
 Data File : VN087124.D
 Acq On : 20 Jun 2025 09:46
 Operator : JC\MD
 Sample : VN0620WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0620WBL01

Quant Time: Jun 20 22:59:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	186699	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	379621	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	343369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	159510	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	137319	54.934	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.860%
35) Dibromofluoromethane	8.171	113	115842	51.492	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.980%
50) Toluene-d8	10.565	98	468951	52.655	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.320%
62) 4-Bromofluorobenzene	12.847	95	161930	48.938	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062025\
Data File : VN087124.D
Acq On : 20 Jun 2025 09:46
Operator : JC\MD
Sample : VN0620WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0620WBL01

Quant Time: Jun 20 22:59:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

