

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062025\
 Data File : VN087132.D
 Acq On : 20 Jun 2025 12:44
 Operator : JC\MD
 Sample : Q2362-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EP-6

Quant Time: Jun 20 23:04:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	195844	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	388591	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	352130	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	166973	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	136410	52.023	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.040%	
35) Dibromofluoromethane	8.171	113	118252	51.350	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.700%	
50) Toluene-d8	10.565	98	479750	52.624	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.240%	
62) 4-Bromofluorobenzene	12.847	95	171364	50.594	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.180%	
Target Compounds						
					Qvalue	
16) Acetone	4.447	43	28355	20.391	ug/l	98
18) Methyl Acetate	5.053	43	24697	6.095	ug/l	96
20) Methylene Chloride	5.295	84	3998	1.536	ug/l #	92
37) Ethyl Acetate	7.565	43	16984	3.888	ug/l	98
43) Isopropyl Acetate	8.694	43	10944	1.561	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062025\
Data File : VN087132.D
Acq On : 20 Jun 2025 12:44
Operator : JC\MD
Sample : Q2362-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP-6

Quant Time: Jun 20 23:04:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

