

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062025\
 Data File : VN087133.D
 Acq On : 20 Jun 2025 13:04
 Operator : JC\MD
 Sample : Q2367-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EP-4

Quant Time: Jun 20 23:04:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	217895	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	434510	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	395910	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	186922	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	152750	52.359	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.720%	
35) Dibromofluoromethane	8.171	113	131339	51.005	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.020%	
50) Toluene-d8	10.565	98	536753	52.655	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.300%	
62) 4-Bromofluorobenzene	12.847	95	188760	49.840	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 99.680%	
Target Compounds						Qvalue
16) Acetone	4.447	43	28691	18.545	ug/l	98
18) Methyl Acetate	5.047	43	25608	5.680	ug/l	98
20) Methylene Chloride	5.294	84	4758	1.643	ug/l #	79
37) Ethyl Acetate	7.571	43	16383	3.354	ug/l	99
43) Isopropyl Acetate	8.688	43	10735	1.369	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062025\
Data File : VN087133.D
Acq On : 20 Jun 2025 13:04
Operator : JC\MD
Sample : Q2367-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP-4

Quant Time: Jun 20 23:04:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

