

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062025\
 Data File : VN087137.D
 Acq On : 20 Jun 2025 14:27
 Operator : JC\MD
 Sample : Q2377-01 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PW-B6-L66-061925

Quant Time: Jun 20 23:06:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	187295	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	394752	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	381540	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	178554	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	142816	56.952	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.900%
35) Dibromofluoromethane	8.177	113	118312	50.574	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.140%
50) Toluene-d8	10.565	98	467107	50.438	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.880%
62) 4-Bromofluorobenzene	12.847	95	182384	53.007	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.020%
Target Compounds						
30) Chloroform	7.977	83	1286	0.307	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062025\
Data File : VN087137.D
Acq On : 20 Jun 2025 14:27
Operator : JC\MD
Sample : Q2377-01 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PW-B6-L66-061925

Quant Time: Jun 20 23:06:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

