

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062025\
 Data File : VN087141.D
 Acq On : 20 Jun 2025 15:50
 Operator : JC\MD
 Sample : Q2355-01
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 3897

Quant Time: Jun 20 23:08:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	180643	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	9.100	114	377941	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	326455	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	154668	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	130607	54.001	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.000%
35) Dibromofluoromethane	8.171	113	104571	46.688	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.380%
50) Toluene-d8	10.565	98	427998	48.271	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.540%
62) 4-Bromofluorobenzene	12.847	95	150793	45.775	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.560%
Target Compounds						
						Qvalue
10) Methyl Iodide	4.594	142	6171	2.417	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062025\
Data File : VN087141.D
Acq On : 20 Jun 2025 15:50
Operator : JC\MD
Sample : Q2355-01
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
3897

Quant Time: Jun 20 23:08:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

