

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062025\
 Data File : VN087136.D
 Acq On : 20 Jun 2025 14:06
 Operator : JC\MD
 Sample : Q2377-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB01-061925

Quant Time: Jun 20 23:05:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	168949	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	357590	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	334632	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	157871	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	125141	55.322	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.640%
35) Dibromofluoromethane	8.171	113	101496	47.894	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.780%
50) Toluene-d8	10.565	98	414042	49.354	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.700%
62) 4-Bromofluorobenzene	12.847	95	160931	51.633	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.260%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062025\
Data File : VN087136.D
Acq On : 20 Jun 2025 14:06
Operator : JC\MD
Sample : Q2377-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB01-061925

Quant Time: Jun 20 23:05:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1068
Delta R.T. -0.006 min
Lab File: VN087136.D
Acq: 20 Jun 2025 14:06




Instrument :
MSVOA_N
ClientSampleId :
TB01-061925


Tgt Ion:168 Resp: 168949
Ion Ratio Lower Upper
168 100
99 59.4 49.1 73.7



#33
1,2-Dichloroethane-d4
Concen: 55.322 ug/l
RT: 8.583 min Scan# 1127
Delta R.T. -0.006 min
Lab File: VN087136.D
Acq: 20 Jun 2025 14:06

Tgt Ion: 65 Resp: 125141
Ion Ratio Lower Upper
65 100
67 54.6 0.0 105.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. -0.000 min

Lab File: VN087136.D

Acq: 20 Jun 2025 14:06

Instrument :

MSVOA_N

ClientSampleId :

TB01-061925





Tgt Ion:114 Resp: 357590

Ion Ratio Lower Upper

114 100

63 18.5 0.0 39.6

88 15.6 0.0 30.2



#35

Dibromofluoromethane

Concen: 47.894 ug/l

RT: 8.171 min Scan# 1057

Delta R.T. -0.000 min

Lab File: VN087136.D

Acq: 20 Jun 2025 14:06




Tgt Ion:113 Resp: 101496


Ion Ratio Lower Upper

113 100

111 102.9 84.2 126.2

192 18.0 14.2 21.4









#50
Toluene-d8
Concen: 49.354 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. -0.006 min
Lab File: VN087136.D
Acq: 20 Jun 2025 14:06




Instrument :
MSVOA_N
ClientSampleId :
TB01-061925


Tgt Ion: 98 Resp: 414042
Ion Ratio Lower Upper
98 100
100 66.2 53.4 80.0



#62
4-Bromofluorobenzene
Concen: 51.633 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.000 min
Lab File: VN087136.D
Acq: 20 Jun 2025 14:06

Tgt Ion: 95 Resp: 160931
Ion Ratio Lower Upper
95 100
174 73.2 0.0 141.8
176 70.0 0.0 132.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN087136.D
Acq: 20 Jun 2025 14:06

Instrument :
MSVOA_N
ClientSampleId :
TB01-061925

Tgt Ion:117 Resp: 334632
Ion Ratio Lower Upper
117 100
82 55.1 44.6 67.0
119 32.4 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.006 min
Lab File: VN087136.D
Acq: 20 Jun 2025 14:06

Tgt Ion:152 Resp: 157871
Ion Ratio Lower Upper
152 100
115 59.3 30.1 90.5
150 158.5 0.0 345.0

