

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062119\
 Data File : VN056432.D
 Acq On : 21 Jun 2019 17:36
 Operator : JC/SP
 Sample : PB120805ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB120805ZHE#03

Quant Time: Jun 22 03:31:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	388407	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	652406	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	644541	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	195312	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	221434	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
35) Dibromofluoromethane	7.59	113	192570	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
50) Toluene-d8	10.09	98	809201	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
62) 4-Bromofluorobenzene	12.40	95	266524	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
Target Compounds						
16) Acetone	3.82	43	21173	9.817	ug/l	95
20) Methylene Chloride	4.55	84	12049	2.111	ug/l	98
43) Isopropyl Acetate	8.17	43	17673	2.031	ug/l #	91
95) Naphthalene	15.13	128	4785	5.724	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062119\
Data File : VN056432.D
Acq On : 21 Jun 2019 17:36
Operator : JC/SP
Sample : PB120805ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB120805ZHE#03

Quant Time: Jun 22 03:31:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

