

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062119\
 Data File : VN056433.D
 Acq On : 21 Jun 2019 18:00
 Operator : JC/SP
 Sample : PB120805ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB120805ZHE#04

Quant Time: Jun 22 03:32:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	404065	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	687853	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	688406	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	215492	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	226216	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
35) Dibromofluoromethane	7.59	113	200410	46.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.40%	
50) Toluene-d8	10.09	98	860502	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
62) 4-Bromofluorobenzene	12.40	95	286246	51.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.86%	
Target Compounds						
16) Acetone	3.82	43	19739	8.798	ug/l #	86
20) Methylene Chloride	4.55	84	12391	2.076	ug/l	97
43) Isopropyl Acetate	8.17	43	17319	1.888	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062119\
Data File : VN056433.D
Acq On : 21 Jun 2019 18:00
Operator : JC/SP
Sample : PB120805ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB120805ZHE#04

Quant Time: Jun 22 03:32:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

