

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062119\
 Data File : VN056435.D
 Acq On : 21 Jun 2019 18:47
 Operator : JC/SP
 Sample : K3158-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-03-062019

Quant Time: Jun 22 03:40:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	381707	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	647777	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	627938	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	190396	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	216909	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
35) Dibromofluoromethane	7.59	113	187012	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
50) Toluene-d8	10.09	98	798351	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
62) 4-Bromofluorobenzene	12.40	95	258457	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
Target Compounds						
20) Methylene Chloride	4.55	84	176359	44.032	ug/l	98
43) Isopropyl Acetate	8.17	43	15004	1.737	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062119\
Data File : VN056435.D
Acq On : 21 Jun 2019 18:47
Operator : JC/SP
Sample : K3158-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-03-062019

Quant Time: Jun 22 03:40:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

