

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062121\
 Data File : VN067407.D
 Acq On : 21 Jun 2021 11:02
 Operator : JC/MD
 Sample : VN0621MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0621MBL01

Quant Time: Jun 22 03:09:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	299296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	589712	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	520238	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	182492	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	286593	48.540	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.080%
35) Dibromofluoromethane	8.024	113	184125	46.946	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.900%
50) Toluene-d8	10.443	98	777733	51.737	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.480%
62) 4-Bromofluorobenzene	12.736	95	273423	44.074	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062121\
Data File : VN067407.D
Acq On : 21 Jun 2021 11:02
Operator : JC/MD
Sample : VN0621MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0621MBL01

Quant Time: Jun 22 03:09:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

