

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062121\
 Data File : VN067420.D
 Acq On : 21 Jun 2021 16:38
 Operator : JC/MD
 Sample : M2795-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 B1GW

Manual Integrations
 APPROVED

MMDadoda
 6/22/2021 6:25:18 PM

Quant Time: Jun 22 03:11:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	286683	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	584834	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	515670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	169031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	282862	50.016	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.040%			
35) Dibromofluoromethane	8.024	113	181277	46.606	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 93.220%			
50) Toluene-d8	10.446	98	754032	50.578	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 101.160%			
62) 4-Bromofluorobenzene	12.733	95	259551	42.187	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 84.380%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	6178m	2.847	ug/l	
17) Carbon Disulfide	4.524	76	5882	0.530	ug/l	95
39) Methylcyclohexane	9.459	83	4430	0.638	ug/l #	75
65) Chlorobenzene	11.776	112	4612	0.411	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062121\
Data File : VN067420.D
Acq On : 21 Jun 2021 16:38
Operator : JC/MD
Sample : M2795-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
B1GW

Manual Integrations
APPROVED

MMDadoda
6/22/2021 6:25:18 PM

Quant Time: Jun 22 03:11:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

