

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062121\
 Data File : VN067430.D
 Acq On : 21 Jun 2021 20:41
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 22 03:13:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	73	0.00
2 T	Dichlorodifluoromethane	0.734	0.886	-20.7	81	0.00
3 P	Chloromethane	0.759	0.925	-21.9	92	0.00
4 C	Vinyl Chloride	0.744	0.894	-20.2#	83	0.00
5 T	Bromomethane	0.316	0.444	-40.5#	99	0.00
6 T	Chloroethane	0.457	0.566	-23.9	90	0.00
7 T	Trichlorofluoromethane	1.033	1.275	-23.4	89	0.00
8 T	Diethyl Ether	0.420	0.489	-16.4	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.621	0.758	-22.1	87	0.00
10 T	Methyl Iodide	0.546	0.789	-44.5#	101	0.00
11 T	Tert butyl alcohol	0.162	0.188	-16.0	88	0.00
12 CM	1,1-Dichloroethene	0.600	0.751	-25.2#	90	0.00
13 T	Acrolein	0.101	0.128	-26.7#	96	0.00
14 T	Allyl chloride	1.331	1.558	-17.1	86	0.00
15 T	Acrylonitrile	0.356	0.495	-39.0#	96	0.00
16 T	Acetone	0.379	0.421	-11.1	87	0.00
17 T	Carbon Disulfide	1.937	2.280	-17.7	88	0.00
18 T	Methyl Acetate	0.918	1.178	-28.3#	97	0.00
19 T	Methyl tert-butyl Ether	2.407	3.014	-25.2#	86	0.00
20 T	Methylene Chloride	0.788	0.928	-17.8	93	0.00
21 T	trans-1,2-Dichloroethene	0.665	0.783	-17.7	85	0.00
22 T	Diisopropyl ether	2.456	3.216	-30.9#	92	0.00
23 T	Vinyl Acetate	2.250	2.823	-25.5#	89	0.00
24 P	1,1-Dichloroethane	1.378	1.739	-26.2#	90	0.00
25 T	2-Butanone	0.541	0.684	-26.4#	91	0.00
26 T	2,2-Dichloropropane	1.396	1.324	5.2	70	0.00
27 T	cis-1,2-Dichloroethene	0.782	0.926	-18.4	87	0.00
28 T	Bromochloromethane	0.641	0.789	-23.1	93	0.00
29 T	Tetrahydrofuran	0.341	0.448	-31.4#	92	0.00
30 C	Chloroform	1.446	1.706	-18.0#	86	0.00
31 T	Cyclohexane	1.341	1.571	-17.2	89	0.00
32 T	1,1,1-Trichloroethane	1.329	1.466	-10.3	78	0.00
33 S	1,2-Dichloroethane-d4	0.986	1.055	-7.0	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.333	0.322	3.3	83	0.00
36 T	1,1-Dichloropropene	0.561	0.598	-6.6	87	0.00
37 T	Ethyl Acetate	0.594	0.662	-11.4	91	0.00
38 T	Carbon Tetrachloride	0.603	0.556	7.8	75	0.00
39 T	Methylcyclohexane	0.593	0.684	-15.3	90	0.00
40 TM	Benzene	1.565	1.758	-12.3	91	0.00
41 T	Methacrylonitrile	0.309	0.331	-7.1	88	0.00
42 TM	1,2-Dichloroethane	0.661	0.674	-2.0	82	0.00
43 T	Isopropyl Acetate	1.066	1.156	-8.4	88	0.00
44 TM	Trichloroethene	0.366	0.363	0.8	80	0.00
45 C	1,2-Dichloropropane	0.417	0.486	-16.5#	94	0.00
46 T	Dibromomethane	0.284	0.300	-5.6	85	0.00
47 T	Bromodichloromethane	0.637	0.650	-2.0	84	0.00
48 T	Methyl methacrylate	0.486	0.511	-5.1	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062121\
 Data File : VN067430.D
 Acq On : 21 Jun 2021 20:41
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 22 03:13:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.008	-14.3	94	0.00
50 S	Toluene-d8	1.275	1.284	-0.7	85	0.00
51 T	4-Methyl-2-Pentanone	0.614	0.694	-13.0	91	0.00
52 CM	Toluene	0.984	1.046	-6.3#	86	0.00
53 T	t-1,3-Dichloropropene	0.705	0.722	-2.4	82	0.00
54 T	cis-1,3-Dichloropropene	0.713	0.764	-7.2	87	0.00
55 T	1,1,2-Trichloroethane	0.380	0.407	-7.1	88	0.00
56 T	Ethyl methacrylate	0.684	0.756	-10.5	90	0.00
57 T	1,3-Dichloropropane	0.695	0.777	-11.8	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.199	0.203	-2.0	86	0.00
59 T	2-Hexanone	0.450	0.502	-11.6	90	0.00
60 T	Dibromochloromethane	0.416	0.405	2.6	78	0.00
61 T	1,2-Dibromoethane	0.398	0.406	-2.0	84	0.00
62 S	4-Bromofluorobenzene	0.526	0.480	8.7	78	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.355	0.337	5.1	75	0.00
65 PM	Chlorobenzene	1.088	1.133	-4.1	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.405	0.404	0.2	78	0.00
67 C	Ethyl Benzene	2.152	2.303	-7.0#	84	0.00
68 T	m/p-Xylenes	0.765	0.789	-3.1	82	0.00
69 T	o-Xylene	0.748	0.788	-5.3	83	0.00
70 T	Styrene	1.244	1.299	-4.4	80	0.00
71 P	Bromoform	0.310	0.285	8.1	72	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	75	0.00
73 T	Isopropylbenzene	4.798	5.300	-10.5	82	0.00
74 T	N-amyl acetate	2.342	2.811	-20.0	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.430	1.779	-24.4	93	0.00
76 T	1,2,3-Trichloropropane	1.298	1.493	-15.0	84	0.00
77 T	Bromobenzene	0.977	1.002	-2.6	77	0.00
78 T	n-propylbenzene	5.706	6.491	-13.8	84	0.00
79 T	2-Chlorotoluene	3.553	3.938	-10.8	83	0.00
80 T	1,3,5-Trimethylbenzene	4.042	4.391	-8.6	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.524	0.581	-10.9	84	0.00
82 T	4-Chlorotoluene	3.538	3.849	-8.8	80	0.00
83 T	tert-Butylbenzene	3.218	3.540	-10.0	81	0.00
84 T	1,2,4-Trimethylbenzene	3.993	4.278	-7.1	79	0.00
85 T	sec-Butylbenzene	4.556	5.163	-13.3	83	0.00
86 T	p-Isopropyltoluene	3.677	3.947	-7.3	80	0.00
87 T	1,3-Dichlorobenzene	1.762	1.822	-3.4	75	0.00
88 T	1,4-Dichlorobenzene	1.773	1.802	-1.6	75	0.00
89 T	n-Butylbenzene	3.455	3.884	-12.4	83	0.00
90 T	Hexachloroethane	0.719	0.808	-12.4	80	0.00
91 T	1,2-Dichlorobenzene	1.672	1.797	-7.5	77	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.358	0.363	-1.4	76	0.00
93 T	1,2,4-Trichlorobenzene	0.873	0.853	2.3	70	0.00
94 T	Hexachlorobutadiene	0.501	0.417	16.8	62	0.00
95 T	Naphthalene	2.660	2.810	-5.6	73	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062121\
Data File : VN067430.D
Acq On : 21 Jun 2021 20:41
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 22 03:13:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.838	0.834	0.5	72	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6