

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
 Data File : VN072971.D
 Acq On : 21 Jun 2022 09:55
 Operator : JC\MD
 Sample : VN0621MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0621MBL01

Quant Time: Jun 22 06:58:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	337773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	538073	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.674	117	491982	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.604	152	240664	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	195439	42.797	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.600%
35) Dibromofluoromethane	7.951	113	172106	51.749	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.500%
50) Toluene-d8	10.375	98	550808	44.509	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.020%
62) 4-Bromofluorobenzene	12.663	95	243843	54.796	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062122\
Data File : VN072971.D
Acq On : 21 Jun 2022 09:55
Operator : JC\MD
Sample : VN0621MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0621MBL01

Quant Time: Jun 22 06:58:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

